


	<p><b>UN38.3 Lithium Battery Test Summary for PyroScience GmbH</b> Item Nr. (Product name):</p> <ul style="list-style-type: none"> <li>FSGO2 (FireSting-GO2)</li> <li>APHOX-TX, APHOX-L-O2, APHOX-L-PH (AquapHOx Logger)</li> </ul>
1	<p><b>Battery Manufacturer</b> LiPol Battery Co., Ltd.</p>
2	<p><b>Product Manufacturer</b> PyroScience GmbH Kackertstr. 11 52072 Aachen, Germany <a href="mailto:info@pyroscience.com">info@pyroscience.com</a> +49 241 518 322 10</p>
3	<p><b>UN38.3 Testing Laboratory</b> KSIGN(Guangdong) Testing Co., Ltd. West Side of 1/F., Building C, Zone A, Fuyuan New Factory, Jiujiu Industrial Park, Minzhu, Shatou, Shajing, Bao'an District, Shenzhen, Guangdong, China</p>
4	<p><b>Test Report Number</b> KS2302S0659B01</p>
5	<p><b>Date of Test Report</b> 03 March 2023</p>
6	<p><b>Description of Cell or Battery</b> 3.7V, 1500mAh, 5.55Wh, 28g, lithium polymer battery (single cell battery), Modell LP753450</p>
7	<p><b>UN38.3 Tests Performed and Successfully Passed</b> Altitude simulation, thermal test, vibration, shock, external short-circuit, crush, overcharge, forced discharge</p>
8	<p><b>Assembled Battery Testing Requirements</b> Not Applicable</p>
9	<p><b>Edition of UN Manual of Tests and Criteria Used</b> UN Manual of Tests and Criteria ST/SG/AC.10/11/Rev.7/Amend.1 38.3</p>
10	<p><b>Name and Title of Signatory</b></p> <p> Dr. Roland Thar, PyroScience Engineer</p>

  
 PyroScience GmbH  
 Kackertstr. 11 +49 241 5183 2210  
 DE-52072 Aachen www.pyroscience.com  
 info@pyroscience.com